



VVTD0201XX




2HP LUBE VANE TANK OVER ASSEMBLY

REVISION NOTES		
DESCRIPTION	ECN	PR/NOR
UPDATED FLOW DIAGRAM	PXEC0388	N/A

POWEREX™



30GAL ASME RECEIVER
(Tanque ASME de 114 litros)




- NOTES:
- APPROXIMATE WEIGHT: 595 lbm
 - A MINIMUM OF 24" CLEARANCE IS RECOMMENDED AROUND THE ENTIRE UNIT. CLEARANCE FOR CONTROL PANEL SHALL BE IN ACCORDANCE WITH OSHA/NEC REQUIREMENTS.
 - SYSTEM DESIGNED IN COMPLIANCE WITH U.S. NFPA '99, LATEST EDITION.
- PESO APROXIMADO: 270 kg
 - RECOMENDAMOS QUE EL ESPACIO LIBRE ALREDEDOR DEL SISTEMA DEBE SER COMO MÍNIMO 24" (600 mm). EL ESPACIO LIBRE PARA EL TABLERO DE CONTROL DEBERÁ ESTAR DE ACUERDO CON LOS REQUERIMIENTOS DE SEGURIDAD DE LA OSHA/NEC.
 - EL SISTEMA ESTÁ DISEÑADO DE ACUERDO CON LA ÚLTIMA EDICIÓN DE LA NFPA 99.

SYMBOL LEGEND		DRAWINGS TO CONFORM TO CURRENT REVISION OF THE FOLLOWING ANSI STANDARDS: Y14.3 MULTI AND SECTIONAL VIEW DWGS Y14.5 DIMENSIONING AND TOLERANCING	
TOLERANCES, UNLESS OTHERWISE SPECIFIED:	SOLID EDGE DRAWING	DO NOT SCALE DRAWING FOR DIMENSIONS	
0.XXX ± .005	INCH [mm]	SHEET 1 OF 2	SCALE 1:8
0.XXX ± .005			
CONFIDENTIAL DISCLOSURE This drawing is the property of POWEREX of the SCOTT FEZTER COMPANY and is subject to return on demand. Its contents are confidential and must not be copied or submitted to outside parties for use or examination.		AUTHOR DWright	REVISION DATE 6/29/2021
ITEM NO: 729923	POWEREX 150 PRODUCTION DRIVE, HARRISON OH 45030 (888) 769-7929		
DESCRIPTION 2HP LUBE VANE TANK OVER ASSEMBLY	DRAWING NUMBER VVTD0201XX	REVISION D	SIZE D

VVTD0201XX

2HP LUBE VANE TANK OVER ASSEMBLY

REVISION NOTES		
DESCRIPTION	ECN	PR/NOR
UPDATED FLOW DIAGRAM	PXEC0388	N/A



POWEREX™

LEGEND (Leyenda)	
	CHECK VALVE (Válvula Cheque)
	BALL VALVE (Válvula de Bola)
	INTAKE FILTER (Filtro admisión)
	GAUGE (Manómetro)
	UNION (Unión)
	DRAIN COCK (Grifo de Drenaje)
	AIR LINE (Tubería de Aire)
	VACUUM PUMP (Bomba de Succión)
	BUTTERFLY VALVE (Válvula de Mariposa)
	FLEX HOSE (Manguera flexible)
	TEMPERTURE SWITCH (Interruptor temperatura)
	VACUUM SWITCH/TRANSDUCER (Interruptor/Transductor vacío)
	SOLENOID VALVE (Válvula solenoide)
	CONDENSATE LINE (Línea de Condensado)

NOTES:
1) PER NFPA THE SYSTEM WILL REQUIRE A SOURCE VALVE (SUPPLIED BY OTHERS)

SYMBOL LEGEND		DRAWINGS TO CONFORM TO CURRENT REVISION OF THE FOLLOWING AISC STANDARDS: Y14.1 MULTI AND SECTIONAL VIEW DWGS Y14.5 DIMENSIONING AND TOLERANCING	
TOLERANCES, UNLESS OTHERWISE SPECIFIED	SOLID EDGE DRAWING	DO NOT SCALE DRAWING FOR DIMENSIONS	
0.XXX ± .005	INCH [mm]	SHEET 2 OF 2	SCALE 1:1
0.000 ± .005			
CONFIDENTIAL DISCLOSURE This drawing is the property of POWEREX of the SCOTT FEZTER COMPANY and is subject to return on demand. Its contents are confidential and must not be copied or submitted to outside parties for use or examination.		AUTHOR DWright	REVISION DATE 6/29/2021
ITEM NO: 729923		APPROVALS ON FILE IN PLM	
MATERIAL: ---		POWEREX 150 PRODUCTION DRIVE, HARRISON OH 45030 (888) 769-7979	
DESCRIPTION 2HP LUBE VANE TANK OVER ASSEMBLY	DRAWING NUMBER VVTD0201XX	REVISION D	SIZE D